

Technical Data

	Unit	LSA	
Measurement Range	Standard (330 – 1180 nm)	■	
	UV-I (248 – 1180 nm)	■	
	UV-II (192 – 800 nm)	■	
	UV-II-VIS (192 – 1180 nm)	■	
	VIS/IR (330 – 1750 nm)	■	
	IR-I (630 – 1750 nm)	■	
	IR-II (1000 – 2250 nm)	■	
	IR-III (1400 – 11000 nm)	■ ¹⁾	
Absolute Accuracy ²⁾	192 – 330 nm ³⁾	pm	6
	330 – 420 nm	pm	3
	420 – 1100 nm		6
	IR-I	GHz	12
	IR-II		25
	IR-III	nm	1 – 5 ¹⁾
Quick Coupling Accuracy (with multi mode fiber)	GHz	20 ⁴⁾	
Wavelength Deviation Sensitivity/ Measurement Resolution	192 – 330 nm ³⁾	pm	5
	330 – 420 nm	pm	2
	420 – 1100 nm		3
	IR-I	GHz	6
	IR-II		12
	IR-III	nm	1
Resolving Power ($\lambda/\Delta\lambda$) ⁵⁾	Standard / UV	Singlemode Multimode fiber	20000 10000
	IR-I		4000 2000
	IR-II		2800 2000
	IR-III		15 – 30 nm ¹⁾
	Standard / UV		7
Linewidth Measurement Accuracy ⁶⁾	IR-I	GHz	40
	IR-II		60
	IR-III		15 % (≥ 200 GHz)
	Standard / UV		7
Maximal Linewidth	THz	1.5	
Measurement Speed ⁷⁾	Data Acquisition		500
	Wavelength Calculation	Hz	60
	Spectrum Calculation		15
Required Input Energy and Power ⁸⁾	Standard	μ J	0.0001 – 0.04
	UV-I, UV-II	(or μ W)	0.0001 – 0.1
	IR-I, IR-II		0.02 – 2
	IR-III	mW	1 ¹⁾
Diffraction Grating	FSR	THz	~5.4
Coupling Fiber Diameter			50 μ m or single mode fiber set
Calibration			Built-in calibration ⁹⁾
Calibration Period			≤ 1 month
Warm-up Time			No warm-up time under constant ambient conditions. Otherwise until thermal and air pressure equilibrium is reached
Dimensions L x W x H	mm		325 x 180 x 77
Weight	kg		2.8
Interface			High-speed USB 2.0 connection
Power Supply			Power consumption < 2.3 W, supply directly via USB cable; IR-II & IR-III: external power supply included

1) For further information on IR-III devices see upper table on following page 2) According to 3 σ criterion 3) With multi mode fiber
4) Only for standard range 5) Spectral resolution $\Delta\lambda = \lambda / R$; R = resolving power. According to Rayleigh criterion. 6) But not better than 5% of the linewidth
7) Depending on PC hardware and settings. Without autocalibration usage

Technical Data

	Unit	LSA IR-III TYPE 2 – 3	LSA IR-III TYPE 2 – 6	LSA IR-III TYPE 2 – 11
Measurement Range	nm	1400 – 3000	1400 – 6000	1400 – 11000
Absolute Accuracy ²⁾	nm	1	2	5
Relative Accuracy		1.25×10^{-4}	3×10^{-4}	5×10^{-4}
Wavelength Deviation Sensitivity/Measurement Resolution		0.7×10^{-4}	1.5×10^{-4}	2.5×10^{-4}
Resolving Power ($\lambda/\Delta\lambda$) ⁵⁾	nm	15	20	30
Linewidth Measurement Accuracy ⁶⁾			15%	
Maximal Linewidth	THz		1 (up to 15)	
Measurement Speed ⁷⁾	Data Acquisition		100	
	Wavelength Calculation	Hz	100	
	Spectrum Calculation		15	
Required Input Energy and Power ⁸⁾	Pulsed	μ J	10	
	cw	mW	0.2	
Diffraction Grating	FSR	THz	~2.7	
Coupling Fiber			PIR-550/600 or CIR-550/600	
Calibration			SLR-1532 or 3.39 μ m HeNe calibration laser (not included)	
Calibration Period			≤ 15 days	
Warm-up Time			No warm-up time under constant ambient conditions. Otherwise until thermal and air pressure equilibrium is reached	
Dimensions L x W x H	mm		325 x 180 x 77	
Weight	kg		3.0	
Interface			High-speed USB 2.0 connection	
Power Supply			External power supply included	

Technical Data

	Unit	HDSA UV-II	HDSA UV-I	HDSA Standard	HDSA IR-I	HDSA Telecom ¹¹⁾
Measurement Range	nm	192 – 400	330 – 800	350 – 1050	940 – 1740	1500 – 1600
Absolute Accuracy ²⁾	GHz	20	5	5	20	3
Wavelength Deviation Sensitivity/ Measurement Resolution	GHz	5	1	2	2	0.6
Resolving Power ($\lambda/\Delta\lambda$) ⁵⁾		10000 @ 325 nm	10000 @ 325 nm	30000 @ 633 nm	5000 @ 1500 nm	20000 @ 1500 nm
	Measure- ment Speed ⁷⁾	Data Acquisition	1	16	7.5	60
	Wavelength Calculation	1	1	1	50	60
	Spectrum Calculation	1	16	7.5	50	60
Required Input Energy and Power ⁸⁾	nJ	0.05 @ 325 nm	0.05 @ 325 nm	0.01 @ 633 nm	50 @ 1500 nm	100 @ 1500 nm
Calibration		External calibration source (included in delivery)				
Calibration Period		≤ 7 days				
Warm-up Time		No warm-up time under constant ambient conditions. Otherwise until thermal and air pressure equilibrium is reached				
Dimensions L x W x H	mm	360 x 210 x 120				
Weight	kg	~4.5				
Interface		1000BASE-T Gigabit Ethernet	USB 3	USB 2.0		
Power supply		External power supply included; Power consumption: 5 W		Directly via USB-cable		

8) The cw power interpretation in [μ W] compares to an exposure of 1s (generally the energy needs to be divided by the exposure time to obtain the required power)
9) IR-III: external calibration sources required, e.g. SLR-1532 10) Broad line versions. For further information please contact: info@highfinesse.com
11) Various modifications available: other spectral range, resolution, accuracy and measurement speed. Please contact us for further details!



HRSA

Customize as you wish